
Midmark M11 - v1.0.5
Total Cycles: 2996

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 180.2 F 3.9
4:00 200.6 F 9.4
6:00 220.4 F 7.0
8:00 232.0 F 11.1
10:00 234.4 F 9.2
12:00 237.0 F 9.3
14:00 250.2 F 16.0
16:00 262.9 F 23.3
18:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.8
2:00 271.8 F 29.0
3:00 272.0 F 29.7
4:00 272.5 F 29.6

Min 271.3 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 234.8 F 0.0
10:00 229.7 F 0.0
15:00 229.0 F 0.0
20:00 226.0 F 0.0
25:00 212.8 F 0.0
30:00 234.8 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

10/11/2023 12:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____