

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3920

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2024 05:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
2:00 183.5 F 3.3  
4:00 193.3 F 8.3  
6:00 212.0 F 10.1  
8:00 227.6 F 9.3  
10:00 234.5 F 9.6  
12:00 242.6 F 13.1  
14:00 243.0 F 12.3  
16:00 255.9 F 19.1  
18:00 267.4 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.5 F 30.0  
2:00 272.2 F 29.4  
3:00 271.2 F 28.9  
4:00 272.0 F 29.8  
  
Min 271.2 F 28.9  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 230.3 F 0.0  
10:00 225.5 F 0.0  
15:00 218.6 F 0.0  
20:00 214.8 F 0.0  
25:00 214.5 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:28  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

09/09/2024 06:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_