
Midmark M11 - v1.0.0
Total Cycles: 5402

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2025 08:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 87.5 F 0.2
4:00 85.6 F 1.4
6:00 124.0 F 3.8
8:00 150.0 F 6.4
10:00 179.5 F 9.4
12:00 199.4 F 13.0
14:00 210.1 F 3.9
16:00 223.7 F 7.0
18:00 231.6 F 10.5
20:00 229.4 F 8.4
22:00 236.2 F 9.8
24:00 237.5 F 9.5
26:00 244.1 F 12.4
28:00 255.4 F 18.3
30:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.6 F 28.4
3:00 272.6 F 29.3
4:00 272.1 F 28.6

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 204.5 F 0.0
10:00 196.5 F 0.0
15:00 195.4 F 0.0
20:00 199.3 F 0.0
25:00 208.1 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

03/10/2025 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____