
Midmark M11 - v1.0.5
Total Cycles: 3548

Name: RITTER2
Sterilizer ID: V2313228
04/23/2024 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 71.7 F 0.2
4:00 82.8 F 1.1
6:00 111.3 F 3.1
8:00 145.1 F 6.1
10:00 168.4 F 9.1
12:00 189.9 F 12.6
14:00 205.9 F 4.5
16:00 218.6 F 8.3
18:00 228.9 F 12.4
20:00 231.5 F 8.4
22:00 242.7 F 13.1
24:00 241.7 F 11.8
26:00 253.7 F 17.9
28:00 264.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.2 F 29.4
2:00 272.1 F 29.8
3:00 271.6 F 29.0
4:00 272.4 F 29.7

Min 271.3 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 216.2 F 0.0
10:00 219.4 F 0.0
15:00 208.8 F 0.0
20:00 215.2 F 0.0
25:00 215.8 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:60
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:08
CYCLE COMPLETE

04/23/2024 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____