
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004395
13:22 20/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:49
STERILIZING 3:53
136.4 C 225kPa 3:53
136.2 C 223kPa 4:23
136.2 C 221kPa 4:53
136.3 C 222kPa 5:23
136.1 C 222kPa 5:53
136.2 C 222kPa 6:23
136.0 C 222kPa 6:53
136.2 C 224kPa 7:23
136.2 C 224kPa 7:23
VENTING 7:23
CYCLE ABORTED 7:41
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:41

Digital Signature #
F3EDE521390E0790
