

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1548

Name: RITTER 1  
Sterilizer ID: V2178198  
01/20/2022 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.4 F 0.0  
2:00 178.4 F 3.7  
4:00 185.5 F 8.2  
6:00 211.1 F 10.0  
8:00 223.7 F 8.0  
10:00 238.2 F 13.1  
12:00 237.6 F 10.2  
14:00 238.3 F 8.1  
16:00 249.8 F 15.1  
18:00 262.0 F 22.1  
20:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.7 F 28.3  
3:00 272.2 F 29.0  
4:00 272.4 F 28.9

Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 215.4 F 0.0  
10:00 226.9 F 0.0  
15:00 230.2 F 0.0  
20:00 229.0 F 0.0  
25:00 229.5 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:06  
CYCLE COMPLETE

01/20/2022 03:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_