
Midmark M11 - v1.0.0
Total Cycles: 3551

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2023 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 194.8 F 3.6
4:00 208.5 F 8.4
6:00 221.3 F 8.6
8:00 231.0 F 9.5
10:00 235.8 F 9.1
12:00 244.1 F 13.0
14:00 243.4 F 11.8
16:00 256.3 F 18.7
18:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.7 F 29.1
4:00 272.0 F 28.5

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 216.7 F 0.0
10:00 232.5 F 0.0
15:00 227.4 F 0.0
20:00 225.7 F 0.0
25:00 224.1 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

07/19/2023 02:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____