
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001497
16:10 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:05
136.4 C 223kPa 3:05
136.1 C 223kPa 3:35
136.4 C 223kPa 4:05
136.1 C 223kPa 4:35
136.4 C 226kPa 5:05
136.4 C 224kPa 5:35
136.4 C 225kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 6:35
VENTING 6:35
AIR DRYING 7:26
CYCLE COMPLETE 9:21

Digital Signature #
8D52AAD4A1E204E3
