
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005166
11:42 28/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:20
136.4 C 222kPa 4:20
136.3 C 223kPa 4:50
136.4 C 224kPa 5:20
136.4 C 223kPa 5:50
136.3 C 219kPa 6:20
136.3 C 223kPa 6:50
136.3 C 220kPa 7:20
136.3 C 221kPa 7:50
136.3 C 221kPa 7:50
VENTING 7:50
AIR DRYING 8:46
CYCLE COMPLETE 10:00

Digital Signature #
80F0FBB377A27D23
