
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001795
8:18 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:37
136.0 C 220kPa 3:37
136.2 C 224kPa 4:07
135.9 C 220kPa 4:37
136.2 C 223kPa 5:07
136.3 C 225kPa 5:37
136.3 C 224kPa 6:07
136.0 C 220kPa 6:37
136.4 C 224kPa 7:07
136.4 C 224kPa 7:07
VENTING 7:07
AIR DRYING 7:57
CYCLE COMPLETE 8:29

Digital Signature #
70323E1D058CA9E3
