
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000929
15:50 11/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:59
STERILIZING 4:25
136.1 C 223kPa 4:25
136.4 C 233kPa 4:55
136.4 C 233kPa 5:25
136.3 C 230kPa 5:55
136.3 C 232kPa 6:25
136.3 C 232kPa 6:55
136.4 C 233kPa 7:25
136.4 C 233kPa 7:55
136.4 C 233kPa 7:55
VENTING 7:55
AIR DRYING 8:50
CYCLE COMPLETE 8:55

Digital Signature #
BA6527754E4D2C5C
