
Midmark M11 - v1.0.5
Total Cycles: 1535

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 178.7 F 3.8
4:00 187.5 F 8.5
6:00 209.1 F 8.9
8:00 225.8 F 9.1
10:00 233.4 F 9.6
12:00 242.7 F 13.4
14:00 243.8 F 13.0
16:00 257.3 F 20.1
18:00 269.2 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.8 F 29.9
2:00 271.8 F 29.2
3:00 272.9 F 30.1
4:00 272.4 F 29.6

Min 271.6 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 214.8 F 0.0
10:00 226.4 F 0.0
15:00 236.9 F 0.0
20:00 232.6 F 0.0
25:00 225.9 F 0.0
30:00 220.4 F 0.0

Drying Complete

FILLING: 1:41
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

09/26/2022 02:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____