

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006820  
10:26 15/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:26  
PRESSURIZING 3:27  
STERILIZING 4:30  
136.0 C 220kPa 4:30  
136.4 C 224kPa 5:00  
136.4 C 222kPa 5:30  
136.4 C 224kPa 6:00  
136.3 C 222kPa 6:30  
136.4 C 221kPa 7:00  
136.4 C 223kPa 7:30  
136.4 C 224kPa 8:00  
136.4 C 224kPa 8:00  
VENTING 8:00  
AIR DRYING 8:52  
CYCLE COMPLETE 10:46

-----  
Digital Signature #  
0DEA55DC71765CCA  
-----