

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1340

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2021 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
2:00 201.1 F 3.2  
4:00 218.3 F 8.0  
6:00 234.1 F 13.0  
8:00 231.8 F 8.5  
10:00 242.0 F 12.1  
12:00 241.0 F 11.6  
14:00 239.8 F 10.3  
16:00 253.2 F 16.9  
18:00 264.6 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.1 F 28.6  
2:00 272.1 F 28.9  
3:00 271.8 F 28.3  
4:00 272.5 F 29.0  
  
Min 271.4 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 206.0 F 0.0  
10:00 209.7 F 0.0  
15:00 206.7 F 0.0  
20:00 205.9 F 0.0  
25:00 207.8 F 0.0  
30:00 205.4 F 0.0  
  
Drying Complete

FILLING: 2:04  
HEATING: 19:55  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

11/08/2021 12:57 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01