

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4059

Name: RITTER2  
Sterilizer ID: V2313228  
10/28/2024 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 71.2 F 0.0  
4:00 73.1 F 0.0  
6:00 94.6 F 0.0  
8:00 135.5 F 0.1  
10:00 169.4 F 0.2  
12:00 188.6 F 0.8  
14:00 201.1 F 3.8  
16:00 220.9 F 6.7  
18:00 224.0 F 10.0  
20:00 236.3 F 10.9  
22:00 235.2 F 9.8  
24:00 239.6 F 10.2  
26:00 244.5 F 12.8  
28:00 243.1 F 11.8  
30:00 254.8 F 17.8  
32:00 265.0 F 24.0  
34:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.0  
2:00 271.6 F 28.4  
3:00 272.6 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 195.1 F 0.0  
10:00 205.0 F 0.0  
15:00 205.3 F 0.0  
20:00 201.3 F 0.0  
25:00 210.4 F 0.0  
30:00 217.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 34:01  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:12:29  
CYCLE COMPLETE

10/28/2024 10:41 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_