

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3491

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2023 04:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.9 F 0.0  
2:00 144.0 F 1.2  
4:00 157.2 F 3.1  
6:00 186.4 F 6.1  
8:00 199.8 F 10.4  
10:00 219.2 F 6.5  
12:00 225.1 F 8.7  
14:00 233.9 F 13.2  
16:00 236.6 F 10.2  
18:00 236.5 F 8.9  
20:00 247.8 F 14.3  
22:00 259.0 F 20.5  
24:00 269.4 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.3 F 29.4  
2:00 272.3 F 29.0  
3:00 272.1 F 29.1  
4:00 272.5 F 29.7

Min 270.9 F 28.4  
Max 273.1 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 216.4 F 0.0  
10:00 241.3 F 0.0  
15:00 236.0 F 0.0  
20:00 226.6 F 0.0  
25:00 229.3 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 25:08  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:03:54  
CYCLE COMPLETE

07/05/2023 05:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_