
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008456
15:34 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:15
136.4 C 222kPa 4:15
136.2 C 223kPa 4:45
136.4 C 219kPa 5:15
136.4 C 222kPa 5:45
136.4 C 222kPa 6:15
136.4 C 225kPa 6:45
136.4 C 223kPa 7:15
136.4 C 225kPa 7:45
136.4 C 225kPa 7:45
VENTING 7:45
AIR DRYING 8:39
CYCLE COMPLETE 8:48

Digital Signature #
681A61AE1D01D9EA
