

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3759

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2023 12:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.4 F 0.0  
2:00 159.3 F 0.0  
4:00 183.4 F 2.7  
6:00 202.3 F 6.1  
8:00 217.9 F 10.4  
10:00 227.1 F 6.4  
12:00 237.1 F 11.0  
14:00 234.3 F 8.2  
16:00 242.6 F 11.2  
18:00 246.6 F 13.3  
20:00 258.7 F 19.8  
22:00 269.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.5 F 28.9  
2:00 272.1 F 28.4  
3:00 271.0 F 28.1  
4:00 272.3 F 29.0

Min 271.0 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 226.9 F 0.0  
10:00 218.6 F 0.0  
15:00 187.7 F 0.0  
20:00 181.4 F 0.0  
25:00 182.6 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:01:42  
CYCLE COMPLETE

09/12/2023 01:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_