
Midmark M11 - v1.0.5
Total Cycles: 2567

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.9 F 0.0
2:00 188.6 F 3.7
4:00 203.4 F 8.9
6:00 222.8 F 8.7
8:00 233.0 F 10.8
10:00 233.5 F 8.6
12:00 239.6 F 10.3
14:00 247.3 F 14.5
16:00 259.8 F 21.4
18:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.9
2:00 272.1 F 29.2
3:00 271.5 F 29.3
4:00 272.8 F 29.8

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 232.1 F 0.0
10:00 234.1 F 0.0
15:00 223.1 F 0.0
20:00 224.1 F 0.0
25:00 227.4 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

06/26/2023 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____