
Midmark M11 - v1.0.5
Total Cycles: 3534

Name: RITTER2
Sterilizer ID: V2313228
04/17/2024 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 80.6 F 0.0
4:00 98.9 F 0.0
6:00 103.8 F 0.3
8:00 146.7 F 2.3
10:00 174.4 F 5.3
12:00 194.2 F 9.2
14:00 206.5 F 13.1
16:00 220.3 F 5.8
18:00 231.2 F 9.6
20:00 237.6 F 10.7
22:00 238.6 F 10.6
24:00 235.4 F 8.1
26:00 249.0 F 14.8
28:00 260.0 F 21.1
30:00 270.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.4 F 29.4
2:00 271.8 F 28.5
3:00 272.5 F 29.2
4:00 271.6 F 28.6

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 219.9 F 0.0
10:00 195.4 F 0.0
15:00 192.5 F 0.0
20:00 188.3 F 0.0
25:00 185.5 F 0.0
30:00 183.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:42
CYCLE COMPLETE

04/17/2024 09:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____