
Midmark M11 - v1.0.0
Total Cycles: 4405

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2024 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.0 F 0.0
2:00 140.9 F 1.3
4:00 145.0 F 4.3
6:00 170.1 F 7.9
8:00 190.8 F 12.2
10:00 217.6 F 4.5
12:00 228.7 F 8.7
14:00 238.4 F 13.2
16:00 237.3 F 9.5
18:00 238.3 F 9.4
20:00 245.7 F 12.9
22:00 257.2 F 19.0
24:00 267.6 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 273.0 F 29.0
2:00 271.7 F 28.7
3:00 272.3 F 28.5
4:00 272.8 F 29.0

Min 270.9 F 28.0
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 237.0 F 0.0
10:00 216.2 F 0.0
15:00 211.0 F 0.0
20:00 207.0 F 0.0
25:00 208.2 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:26
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:03:50
CYCLE COMPLETE

03/26/2024 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____