
Midmark M11 - v1.0.5
Total Cycles: 2945

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 162.8 F 3.5
4:00 183.4 F 7.8
6:00 204.8 F 13.1
8:00 217.0 F 7.5
10:00 230.8 F 12.9
12:00 235.9 F 10.2
14:00 236.1 F 8.5
16:00 249.6 F 15.5
18:00 261.6 F 22.3
20:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.3 F 29.3
2:00 271.8 F 29.4
3:00 272.1 F 29.2
4:00 271.8 F 29.4

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 216.3 F 0.0
10:00 222.8 F 0.0
15:00 214.7 F 0.0
20:00 219.8 F 0.0
25:00 227.8 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

09/28/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____