
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001309
8:45 20/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:13
STERILIZING 5:59
136.4 C 224kPa 5:59
136.4 C 223kPa 6:29
136.4 C 223kPa 6:59
136.2 C 223kPa 7:29
136.4 C 224kPa 7:59
136.4 C 223kPa 8:29
136.2 C 222kPa 8:59
136.4 C 222kPa 9:29
136.4 C 222kPa 9:29
VENTING 9:29
AIR DRYING 10:29
CYCLE COMPLETE 24:11

Digital Signature #
28342BE74A72C12E
