
Midmark M11 - v1.0.5
Total Cycles: 1796

Name: RITTER2
Sterilizer ID: V2313228
11/30/2022 04:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.8 F 0.0
2:00 138.4 F 0.0
4:00 173.2 F 0.4
6:00 199.6 F 3.2
8:00 215.7 F 8.2
10:00 229.9 F 13.3
12:00 234.4 F 10.0
14:00 235.3 F 8.7
16:00 240.7 F 10.8
18:00 248.2 F 14.7
20:00 261.0 F 21.9
22:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.1
2:00 272.5 F 29.6
3:00 271.8 F 28.7
4:00 272.7 F 29.6

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 214.1 F 0.0
10:00 224.2 F 0.0
15:00 230.4 F 0.0
20:00 229.7 F 0.0
25:00 229.1 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

11/30/2022 05:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____