

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3338

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2024 08:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 81.8 F 0.0  
4:00 132.0 F 0.1  
6:00 102.7 F 1.4  
8:00 143.7 F 4.4  
10:00 170.3 F 7.7  
12:00 190.4 F 11.4  
14:00 210.2 F 3.9  
16:00 228.6 F 7.5  
18:00 235.5 F 11.5  
20:00 229.8 F 7.3  
22:00 242.8 F 12.6  
24:00 239.2 F 10.5  
26:00 252.2 F 16.6  
28:00 262.9 F 22.9  
30:00 273.0 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 271.8 F 29.3  
2:00 271.8 F 29.4  
3:00 271.8 F 29.4  
4:00 271.8 F 29.3

Min 271.6 F 29.1  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
5:00 225.9 F 0.0  
10:00 225.3 F 0.0  
15:00 224.2 F 0.0  
20:00 224.6 F 0.0  
25:00 209.3 F 0.0  
30:00 171.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:59  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:59  
CYCLE COMPLETE

01/31/2024 10:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_