

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4170

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2024 03:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.4 F 0.0  
2:00 118.6 F 0.0  
4:00 136.0 F 0.5  
6:00 178.6 F 2.5  
8:00 189.2 F 5.8  
10:00 204.2 F 9.5  
12:00 219.3 F 12.1  
14:00 228.6 F 8.9  
16:00 238.8 F 12.5  
18:00 239.0 F 11.5  
20:00 238.4 F 10.3  
22:00 250.7 F 16.5  
24:00 261.6 F 22.7  
26:00 271.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.7 F 30.1  
2:00 271.9 F 29.4  
3:00 272.0 F 30.0  
4:00 272.6 F 29.9

Min 271.4 F 29.3  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 218.3 F 0.0  
10:00 233.9 F 0.0  
15:00 217.8 F 0.0  
20:00 226.0 F 0.0  
25:00 234.5 F 0.0  
30:00 236.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 26:38  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:05:32  
CYCLE COMPLETE

11/29/2024 04:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_