
Midmark M11 - v1.0.0
Total Cycles: 1225

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2021 06:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 72.6 F 0.0
4:00 82.1 F 0.0
6:00 116.3 F 0.1
8:00 152.5 F 0.2
10:00 182.2 F 0.3
12:00 200.5 F 1.7
14:00 214.9 F 5.1
16:00 222.5 F 8.1
18:00 232.1 F 11.6
20:00 229.8 F 7.0
22:00 242.7 F 12.4
24:00 239.7 F 10.3
26:00 236.0 F 8.5
28:00 251.2 F 15.6
30:00 262.9 F 22.3
32:00 272.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.3
2:00 272.5 F 28.6
3:00 271.8 F 27.9
4:00 272.9 F 28.7

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 202.4 F 0.0
10:00 223.9 F 0.0
15:00 225.8 F 0.0
20:00 216.7 F 0.0
25:00 217.9 F 0.0
30:00 216.8 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:12
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

10/04/2021 07:51 AM
ACCEPTED: X
REJECTED:
Operator: U01