
Midmark M11 - v1.0.0
Total Cycles: 1026

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2021 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	152.9 F	0.0
2:00	148.4 F	0.8
4:00	154.3 F	2.6
6:00	181.8 F	5.9
8:00	201.6 F	10.6
10:00	213.6 F	5.0
12:00	227.3 F	8.3
14:00	240.5 F	13.2
16:00	238.0 F	10.1
18:00	237.9 F	8.6
20:00	248.5 F	14.3
22:00	259.9 F	20.7
24:00	270.4 F	27.4

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.2
1:00	272.7 F	28.7
2:00	271.6 F	28.3
3:00	272.5 F	28.7
4:00	271.5 F	28.2

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

mm:ss	Drying	Start
0:00	203.0 F	0.0
5:00	202.8 F	0.0
10:00	202.3 F	0.0
15:00	208.0 F	0.0
20:00	208.6 F	0.0
25:00	258.0 F	0.0

Dry Cycle Aborted

FILLING: 1:06
HEATING: 24:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 27:59
TOTAL CYCLE: 01:01:07

07/19/2021 09:34 AM
ACCEPTED: X
REJECTED:
Operator: U01