
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000077
8:56 14/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:30
STERILIZING 4:37
136.0 C 220kPa 4:37
136.4 C 225kPa 5:07
136.3 C 221kPa 5:37
136.4 C 225kPa 6:07
136.4 C 222kPa 6:37
136.4 C 223kPa 7:07
136.2 C 223kPa 7:37
136.3 C 223kPa 8:07
136.3 C 223kPa 8:07
VENTING 8:07
AIR DRYING 9:01
CYCLE COMPLETE 24:01

Digital Signature #
CF5DB8055ACA0C59
