

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2642

Name: RITTER2  
Sterilizer ID: V2313228  
07/13/2023 10:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.6 F 0.0  
2:00 169.0 F 1.9  
4:00 166.9 F 5.0  
6:00 195.5 F 8.8  
8:00 205.0 F 13.5  
10:00 220.1 F 7.4  
12:00 232.3 F 12.2  
14:00 233.6 F 9.4  
16:00 238.4 F 10.1  
18:00 245.5 F 13.9  
20:00 256.7 F 20.0  
22:00 267.0 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.5  
1:00 272.0 F 30.2  
2:00 273.0 F 30.8  
3:00 272.5 F 30.3  
4:00 271.7 F 29.9

Min 271.0 F 29.4  
Max 273.3 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 219.6 F 0.0  
10:00 225.6 F 0.0  
15:00 219.4 F 0.0  
20:00 213.7 F 0.0  
25:00 211.4 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:39  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

07/13/2023 11:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_