
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001366
13:10 22/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:41
136.4 C 223kPa 4:41
136.2 C 223kPa 5:11
136.4 C 222kPa 5:41
136.4 C 225kPa 6:11
136.3 C 224kPa 6:41
136.4 C 225kPa 7:11
136.3 C 224kPa 7:41
136.2 C 223kPa 8:11
136.2 C 223kPa 8:11
VENTING 8:11
AIR DRYING 9:03
CYCLE COMPLETE 9:05

Digital Signature #
1978D754C99E41F4
