
Midmark M11 - v1.0.5
Total Cycles: 3645

Name: RITTER2
Sterilizer ID: V2313228
05/24/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 158.7 F 3.6
4:00 178.4 F 7.8
6:00 199.3 F 12.8
8:00 217.2 F 6.2
10:00 230.1 F 11.2
12:00 231.7 F 8.2
14:00 241.9 F 11.6
16:00 244.7 F 13.1
18:00 257.1 F 19.7
20:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.8 F 29.6
2:00 271.7 F 29.0
3:00 272.9 F 29.7
4:00 272.0 F 29.0

Min 271.6 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 206.2 F 0.0
10:00 210.0 F 0.0
15:00 211.1 F 0.0
20:00 206.4 F 0.0
25:00 202.3 F 0.0
30:00 199.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

05/24/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____