
Midmark M11 - v1.0.0
Total Cycles: 4035

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2023 12:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.7 F 0.0
2:00 160.9 F 1.4
4:00 159.3 F 4.2
6:00 170.3 F 7.7
8:00 191.6 F 11.7
10:00 210.4 F 4.1
12:00 218.0 F 8.4
14:00 229.3 F 12.7
16:00 233.8 F 8.9
18:00 240.7 F 11.3
20:00 244.0 F 12.7
22:00 255.9 F 18.7
24:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.3 F 29.4
2:00 271.9 F 28.9
3:00 271.1 F 28.6
4:00 272.0 F 29.4

Min 271.1 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 219.3 F 0.0
10:00 222.2 F 0.0
15:00 230.3 F 0.0
20:00 226.4 F 0.0
25:00 226.2 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 25:42
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

11/17/2023 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____