
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004381
11:22 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:25
STERILIZING 3:11
136.0 C 220kPa 3:11
136.3 C 222kPa 3:41
136.4 C 224kPa 4:11
136.4 C 224kPa 4:41
136.4 C 223kPa 5:11
136.4 C 225kPa 5:41
136.4 C 220kPa 6:11
136.4 C 222kPa 6:41
136.4 C 222kPa 6:41
VENTING 6:41
AIR DRYING 8:45
CYCLE COMPLETE 8:47

Digital Signature #
F8498E8B360697D8
