

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5392

Name: RITTER 1  
Sterilizer ID: V2178198  
03/05/2025 05:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 199.5 F 4.0  
4:00 201.5 F 9.0  
6:00 224.3 F 8.3  
8:00 229.7 F 8.9  
10:00 239.2 F 10.9  
12:00 239.5 F 11.6  
14:00 238.5 F 10.1  
16:00 251.7 F 16.7  
18:00 263.2 F 23.5  
20:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.9  
1:00 272.5 F 29.7  
2:00 272.1 F 29.0  
3:00 272.2 F 29.6  
4:00 272.5 F 29.3

Min 271.5 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 208.6 F 0.0  
10:00 219.1 F 0.0  
15:00 226.0 F 0.0  
20:00 227.5 F 0.0  
25:00 232.1 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

03/05/2025 06:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_