

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2467

Name: RITTER2  
Sterilizer ID: V2313228  
06/05/2023 03:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
2:00 192.4 F 4.4  
4:00 205.1 F 10.0  
6:00 221.5 F 6.4  
8:00 234.6 F 11.9  
10:00 236.3 F 10.1  
12:00 235.6 F 9.0  
14:00 250.9 F 16.4  
16:00 263.3 F 23.6  
18:00 271.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.6  
2:00 271.6 F 29.3  
3:00 272.5 F 29.6  
4:00 271.6 F 29.0  
  
Min 271.5 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 225.9 F 0.0  
10:00 229.6 F 0.0  
15:00 227.4 F 0.0  
20:00 225.7 F 0.0  
25:00 224.4 F 0.0  
30:00 222.5 F 0.0  
  
Drying Complete

FILLING: 2:27  
HEATING: 18:04  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:46  
CYCLE COMPLETE

06/05/2023 04:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_