
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003531
12:36 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:48
136.0 C 220kPa 3:48
136.4 C 223kPa 4:19
136.3 C 224kPa 4:49
136.4 C 225kPa 5:19
136.5 C 225kPa 5:49
136.4 C 226kPa 6:19
136.3 C 223kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:19
VENTING 7:19
AIR DRYING 8:14
CYCLE COMPLETE 8:35

Digital Signature #
D10903D1BD64668D
