
Midmark M11 - v1.0.0
Total Cycles: 4989

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2024 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 74.0 F 0.0
4:00 108.2 F 0.0
6:00 126.6 F 0.2
8:00 153.3 F 0.6
10:00 178.7 F 2.5
12:00 191.9 F 5.5
14:00 202.9 F 8.8
16:00 215.3 F 12.2
18:00 219.7 F 5.0
20:00 230.5 F 9.0
22:00 240.7 F 13.3
24:00 238.7 F 10.2
26:00 236.5 F 8.3
28:00 248.7 F 14.3
30:00 259.9 F 20.5
32:00 270.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.8 F 28.7
2:00 271.6 F 27.8
3:00 272.4 F 28.7
4:00 272.3 F 28.3

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.9 F 0.0
5:00 187.1 F 0.0
10:00 190.8 F 0.0
15:00 191.8 F 0.0
20:00 197.2 F 0.0
25:00 203.5 F 0.0
30:00 207.3 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:58
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:11:05
CYCLE COMPLETE

10/11/2024 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____