

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3128

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2023 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.6 F 0.0  
2:00 147.0 F 1.5  
4:00 156.2 F 4.6  
6:00 183.4 F 8.9  
8:00 202.6 F 11.2  
10:00 220.0 F 7.9  
12:00 232.2 F 13.0  
14:00 235.1 F 10.2  
16:00 235.4 F 8.6  
18:00 248.6 F 15.4  
20:00 260.6 F 22.1  
22:00 271.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.7 F 30.0  
2:00 271.8 F 29.4  
3:00 271.1 F 29.4  
4:00 272.6 F 30.1

Min 271.0 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 210.3 F 0.0  
10:00 210.3 F 0.0  
15:00 211.7 F 0.0  
20:00 210.9 F 0.0  
25:00 215.1 F 0.0  
30:00 217.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

11/15/2023 03:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_