
Midmark M11 - v1.0.0
Total Cycles: 4975

Name: RITTER 1
Sterilizer ID: V2178198
10/08/2024 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 76.2 F 0.0
4:00 106.9 F 0.0
6:00 125.1 F 0.2
8:00 153.2 F 0.8
10:00 175.9 F 3.4
12:00 196.3 F 6.4
14:00 210.6 F 10.4
16:00 216.4 F 7.3
18:00 225.1 F 7.5
20:00 236.4 F 11.9
22:00 233.9 F 8.3
24:00 243.1 F 11.8
26:00 244.3 F 12.3
28:00 256.3 F 18.6
30:00 267.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 28.9
2:00 271.6 F 28.1
3:00 272.4 F 28.9
4:00 272.4 F 28.6

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 203.1 F 0.0
10:00 209.0 F 0.0
15:00 214.2 F 0.0
20:00 224.6 F 0.0
25:00 229.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

10/08/2024 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____