
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000153
7:06 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:33
136.0 C 221kPa 4:33
136.2 C 219kPa 5:03
136.1 C 221kPa 5:33
136.2 C 223kPa 6:03
136.4 C 223kPa 6:33
136.3 C 222kPa 7:03
136.4 C 224kPa 7:33
136.4 C 225kPa 8:03
136.4 C 225kPa 8:03
VENTING 8:03
AIR DRYING 8:54
CYCLE COMPLETE 9:05

Digital Signature #
CB3E6108D83DF4EB
