
Midmark M11 - v1.0.0
Total Cycles: 2723

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2022 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 145.8 F 0.0
4:00 150.3 F 1.6
6:00 174.3 F 4.2
8:00 196.6 F 8.5
10:00 213.9 F 13.0
12:00 222.0 F 7.1
14:00 234.3 F 12.0
16:00 234.0 F 8.7
18:00 240.7 F 10.8
20:00 245.1 F 12.8
22:00 257.1 F 19.2
24:00 267.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.3
2:00 271.8 F 28.6
3:00 271.8 F 28.9
4:00 272.5 F 29.2

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 222.2 F 0.0
10:00 227.9 F 0.0
15:00 229.3 F 0.0
20:00 229.6 F 0.0
25:00 229.0 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

12/28/2022 11:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____