
Midmark M11 - v1.0.5
Total Cycles: 2084

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.3 F 0.0
2:00 166.7 F 1.1
4:00 176.7 F 4.3
6:00 201.4 F 9.5
8:00 224.3 F 7.9
10:00 231.2 F 10.5
12:00 231.7 F 8.0
14:00 240.5 F 10.9
16:00 246.9 F 14.1
18:00 259.6 F 21.2
20:00 270.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.6
2:00 271.2 F 28.9
3:00 272.5 F 29.8
4:00 272.1 F 29.3

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 229.1 F 0.0
10:00 234.0 F 0.0
15:00 224.8 F 0.0
20:00 227.4 F 0.0
25:00 230.9 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

02/23/2023 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____