
Midmark M11 - v1.0.0
Total Cycles: 4740

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2024 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 162.5 F 2.0
4:00 168.0 F 5.4
6:00 189.3 F 9.5
8:00 216.1 F 8.2
10:00 220.7 F 7.2
12:00 235.1 F 12.0
14:00 232.9 F 8.7
16:00 239.7 F 10.6
18:00 244.8 F 13.0
20:00 256.7 F 19.2
22:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.7 F 29.2
2:00 271.2 F 28.5
3:00 272.3 F 29.4
4:00 271.8 F 28.8

Min 271.1 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 207.0 F 0.0
10:00 223.4 F 0.0
15:00 226.2 F 0.0
20:00 228.6 F 0.0
25:00 230.3 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

07/17/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____