

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006519  
11:09 23/10/2023  
RUBBER AND PLASTICS  
121 C for 15 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 3:44  
STERILIZING 4:21  
123.4 C 116kPa 4:21  
123.2 C 118kPa 4:51  
123.4 C 118kPa 5:21  
123.4 C 119kPa 5:51  
123.4 C 119kPa 6:21  
123.4 C 118kPa 6:51  
123.4 C 118kPa 7:21  
123.4 C 118kPa 7:51  
123.4 C 118kPa 8:21  
123.4 C 118kPa 8:51  
123.4 C 118kPa 9:21  
123.4 C 118kPa 9:51  
123.4 C 118kPa 10:21  
123.4 C 118kPa 10:51  
123.3 C 118kPa 11:21  
123.4 C 119kPa 11:51  
123.4 C 118kPa 12:21  
123.4 C 115kPa 12:51  
123.4 C 118kPa 13:21  
123.3 C 115kPa 13:51  
123.4 C 118kPa 14:21  
123.4 C 118kPa 14:51  
123.4 C 118kPa 15:21  
123.4 C 116kPa 15:51  
123.4 C 118kPa 16:21  
123.4 C 119kPa 16:51

123.4 C 116kPa 17:21  
123.4 C 118kPa 17:51  
123.4 C 117kPa 18:21  
123.4 C 119kPa 18:51  
123.4 C 118kPa 19:21  
123.4 C 118kPa 19:21  
VENTING 19:22  
AIR DRYING 20:05  
CYCLE COMPLETE 20:07

-----  
Digital Signature #  
CB98923CF6D9F871  
-----