
Midmark M11 - v1.0.5
Total Cycles: 828

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.5 F 0.0
2:00 199.4 F 4.6
4:00 216.1 F 10.9
6:00 226.7 F 7.7
8:00 238.5 F 11.8
10:00 243.4 F 13.0
12:00 244.5 F 12.9
14:00 258.3 F 20.3
16:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.5
2:00 271.4 F 28.8
3:00 272.5 F 29.7
4:00 272.0 F 29.1

Min 271.3 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 218.4 F 0.0
10:00 233.8 F 0.0
15:00 236.3 F 0.0
20:00 231.7 F 0.0
25:00 227.4 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 2:58
HEATING: 16:49
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:54
CYCLE COMPLETE

02/23/2022 02:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____