
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008117
16:03 28/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:59
STERILIZING 3:04
136.4 C 223kPa 3:04
136.5 C 220kPa 3:34
136.1 C 223kPa 4:04
136.3 C 222kPa 4:34
CYCLE ABORTED 4:38
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:38

Digital Signature #
B92150CC83346B25
