
Midmark M11 - v1.0.5
Total Cycles: 2350

Name: RITTER2
Sterilizer ID: V2313228
05/10/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 188.2 F 3.9
4:00 202.1 F 9.1
6:00 221.1 F 7.9
8:00 232.2 F 10.9
10:00 233.2 F 8.7
12:00 238.9 F 10.1
14:00 247.5 F 14.6
16:00 259.7 F 21.5
18:00 270.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 272.4 F 29.7
2:00 272.2 F 29.9
3:00 271.8 F 29.2
4:00 272.4 F 29.7

Min 271.6 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 223.3 F 0.0
10:00 232.4 F 0.0
15:00 230.0 F 0.0
20:00 222.2 F 0.0
25:00 220.5 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

05/10/2023 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____