

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1097

Name: RITTER 1  
Sterilizer ID: V2178198  
08/19/2021 09:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 168.1 F 2.8  
4:00 191.4 F 7.0  
6:00 214.0 F 13.1  
8:00 224.1 F 7.4  
10:00 238.4 F 12.7  
12:00 238.3 F 10.1  
14:00 235.8 F 8.5  
16:00 251.3 F 15.8  
18:00 263.6 F 22.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.2 F 28.5  
2:00 272.5 F 29.0  
3:00 271.7 F 28.2  
4:00 272.6 F 29.0  
  
Min 271.4 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 197.7 F 0.0  
Dry Cycle Aborted

FILLING: 2:13  
HEATING: 19:59  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 7:15  
TOTAL CYCLE: 00:36:34

08/19/2021 09:47 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02