
Midmark M11 - v1.0.5
Total Cycles: 3664

Name: RITTER2
Sterilizer ID: V2313228
06/03/2024 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.6 F 0.0
2:00 96.4 F 0.4
4:00 113.7 F 2.1
6:00 142.9 F 4.7
8:00 168.3 F 8.1
10:00 184.5 F 11.9
12:00 208.8 F 4.8
14:00 219.0 F 8.7
16:00 227.6 F 13.0
18:00 231.6 F 9.1
20:00 239.2 F 11.2
22:00 242.5 F 12.3
24:00 253.7 F 18.1
26:00 263.8 F 24.2
28:00 272.3 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.6 F 30.1
2:00 271.7 F 29.4
3:00 272.4 F 30.4
4:00 271.9 F 29.8

Min 270.9 F 29.1
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 202.3 F 0.0
10:00 204.8 F 0.0
15:00 213.3 F 0.0
20:00 214.1 F 0.0
25:00 219.3 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:06:35
CYCLE COMPLETE

06/03/2024 04:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____